

XL9 & XL20 Tube Handling Systems



Reliable, Flexible and Affordable

The XL9 & XL20 are reliable, flexible and affordable robotic pick and place systems used to automate critical sample management tasks such as organizing, reformatting and accurately tracking samples. The XL9 and XL20 are fully compatible with LIMS systems and with multiple labware formats. Choose and order your XL Series platform with the capacity and modules right for your lab's application.

applications

- · Compound sample library management
- QA/QC
- · Cherry pick tubes for efficient liquid handling
- Track sample library volume/inventory
- Save freezer space by consolidating tube rack samples
- · Print and apply human readable and barcode labels
- Great benchtop robot for use alongside fully automated sample stores

features

- Re-array, reformat or cherry pick tubes
- Weigh tubes (tare & gross weights)
- 2D Barcode decoding
- LIMS integration ready
- Work list based operation
- · Efficient for both small and large projects
- Label printer and applicator option
- Designed for continuous use and unattended operation
- Encoder based robotics and precision bearings
- provide the highest reliability
- Integration software toolkits available




XL9 and XL20

2D Barcode Scanning

Decode 500 vials per hour

Data Matrix 2D marked sample tubes feature a 2D grid barcode located on the bottom of the tube. This provides a unique ID number for each sample tube. The XL9 and XL20 include built-in 2D scanners that quickly identify and decode the barcode as the tubes are moved from rack to rack. In-process decoding provides an additional level of sample integrity and tracking at the time of tube transfer. The XL Work List Manager software captures and reports the tube 2D barcode, tube location, weight, volume and more to LIMS systems.



Re-array, Reformat, Cherry Pick

Sort up to 700 tubes per hour

For labs with small or large sample libraries, our benchtop tube handlers provide unattended tube processing for work list based tube transfer from source racks to target racks. Labs depend on BioMicroLab's tube sample management systems to increase their liquid handling throughput and to efficiently provide other groups with sample requests.



Label Printing and Application

Print and apply up to 180 vials per hour

Automated tube and vial labeling makes it easy to affix human readable labels to labware and eliminates the headaches associated with manual tube labeling.

- Prints 1D & 2D barcodes and human readable text
- Wide variety of label material and sizes available
- Identical or unique labels
- 300 600 dpi thermal printer
- · User-friendly software
- Available standalone or integrated with the XL200, XL100, XL20 or XL9
- Easily add the LabelPro to your existing XL Series system

Project	Name:								
Open La	ibel C:\Users\User)	(Cesktop/Grochure I	label.xml						
Metroell	Data: C None	C Use merge file	← Use worldist			X	l I ak	elPro	•
						🗥	~ ~	· C	•
Plance File C: Users User 1 Desktop Label Worldst.csv					1843	7C			
✓ Has Header			Generate All Preview Images		LT'#	HE Ass	sav ID:	DNA128	3
Item	Name	Type	Merge		H . T	45 ′ ′ ′ ′ ′ ′ ′ ′ ′ ′ ′ ′ ′ ′ ′ ′ ′ ′ ′	, .D.	D. 17 (112)	
1	BARCODE1	Barcode Text	Bercode Elved						
2	TEXT1 TEXT2	Text	Fixed		-	100 0 \/∢	olume:	50 ul	
3	assay	Text	Assay			_	Jiuiiic.	50 μι	
5	TEXT4	Text	Fixed						
5	volume	Text	Volume			1	:	Г	100
7	TEXTS		Flyad			1.00	arion.	Freezer	A.N
	TEXTS	Text	Fixed				acion.		
9						201	acion.		
	TEXT5 freezer	Text Text	Fixed Location	I Total District	Tables			[tooties	,
	TEXT5 freezer	Text Text	Fixed Location Bercode	Tgt ID	Tgt Pos	Azzay	Volume	Location	
	TEXT5 freezer Src IID W001235446	Text Text Src Pos A01	Fixed Location Bercode ACD124557875	T003342244	A01	Assay DNA120	Volume 30	AZZ	
	TEXT5 freezer Src ID W001235446 W001235446	Text Text Src Pos A01 A02	Fored Location Bercode ACD 124557875 ACD 124557876	T003342244 T003342244	AD1 AD2	Assey DNA120 DNA121	Volume 30 50	AZZ AZ3	
	TEXTS freezer Src ID W001235446 W001235446	Text Text Src Pos A01 A02 A03	Fixed Location Bercode ACD124557875 ACD124557876 ACD124557877	T003342244 T003342244 T003342244	A01	Assey DNA120 DNA121 DNA122	Volume 50 50 50	AZZ AZ3 A24	
9 1 2 3	TEXT5 freezer W001235446 W001235446 W001235446 W001235446	Src Pos A01 A02 A03 A04	Fixed Location Bercode ACD124557873 ACD124557876 ACD124557877 ACD124557878	T003342244 T003342244 T003342244 T003342244	A01 A02 A03 A04	Assey DNA120 DNA121 DNA121 DNA122 DNA122	Volume 50 50 50 50	AZZ AZ3 AZ4 AZ5	
9 1 2 3 4 5	TEXT5 freezer Src ID W001235446 W001235446 W001235446 W001235446	Text Text Src Pos A01 A02 A03 A04 A05	Fixed Location Bercode ACD 124557875 ACD 124557876 ACD 124557877 ACD 124557879 ACD 124557879	T003342244 T003342244 T003342244 T003342244 T003342244	A01 A02 A03 A04 A05	Accesy DNA120 DNA121 DNA122 DNA122 DNA123 DNA124	Volume 50 50 50 50 50 50	AZZ AZ3 AZ4 AZ5 AZ6	
9 1 2 3 4 5 6	TEXT5 freezer W001235446 W001235446 W001235446 W001235446	Src Pos A01 A02 A03 A04	Fixed Location Bercode ACD124557873 ACD124557876 ACD124557877 ACD124557878	T003342244 T003342244 T003342244 T003342244	A01 A02 A03 A04	Assey DNA120 DNA121 DNA121 DNA122 DNA122	Volume 50 50 50 50	AZZ AZ3 AZ4 AZ5	
9 1 2 3 4 5 6 7	TEXT5 freezer W001233446 W001235446 W001235446 W001235446 W001235446 W001235446	Text Text A01 A02 A03 A04 A05 A06	Fixed Location Barcode ACD 124557875 ACD 124557876 ACD 124557878 ACD 124557878 ACD 124557878 ACD 124557879	T003342244 T003342244 T003342244 T003342244 T003342244 T003342244	A01 A02 A03 A04 A05 A06	Areay DNA120 DNA121 DNA122 DNA122 DNA123 DNA124 DNA125	Volume 30 50 50 50 50 50	A22 A23 A24 A25 A26 A27	
9 1 2 3 4 5 6 7 8	TEXT5 freezer W001235446 W001235446 W001235446 W001235446 W001235446 W001235446 W001235446	Text Text Text Src Poe A01 A02 A03 A04 A05 A06 A07	Fixed Location Bercode ACD 124557875 ACD 124557876 ACD 124557876 ACD 124557878 ACD 124557878 ACD 124557881	T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244	A01 A02 A03 A04 A05 A06 A07	ARRHY DNA 120 DNA 121 DNA 122 DNA 122 DNA 123 DNA 124 DNA 125 DNA 126	Volume 30 50 50 50 50 50 50 50 50	A22 A23 A24 A25 A26 A27 A28	
9 1 2 3 4 5 6 7 8	TEXTS freezer Src IID W001235446 W001235446 W001235446 W001235446 W001235446 W001235446 W001235446 W001235446	Text Text A01 A02 A03 A04 A05 A06 A07 A08	Fixed Location Bercode ACD124557875 ACD124557876 ACD124557877 ACD124557879 ACD124557879 ACD124557881 ACD124557881 ACD124557881	T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244	A01 A02 A03 A04 A05 A06 A07 A08	DNA120 DNA121 DNA121 DNA122 DNA123 DNA124 DNA125 DNA126 DNA125	Volume 50 50 50 50 50 50 50 50 50 50	A22 A23 A24 A25 A26 A27 A28 A29	
9 1 2 3 4 5 6 7 8 9	TEXTS freezer Src ID W001235446 W001235446 W001235446 W001235446 W001235446 W001235446 W001235446 W001235446	Text Text Text Text Src Poe A01 A02 A03 A04 A05 A06 A07 A08 A09	Fixed Location Bercode ACD 124557875 ACD 124557876 ACD 124557878 ACD 124557878 ACD 124557878 ACD 124557881 ACD 124557881 ACD 124557882 ACD 124557883	T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244	A01 A02 A03 A04 A05 A06 A07 A08 A09	Assey DNA120 DNA121 DNA122 DNA122 DNA123 DNA124 DNA125 DNA126 DNA126 DNA127 DNA127	Volume 30 50 50 50 50 50 50 50 50 50 5	A22 A23 A24 A25 A26 A27 A28 A29 A30	
9 1 2 3 4 5 6 7 8 9 10 11	TEXTS freezer Src ID W001235446 W001235446 W001235446 W001235446 W001235446 W001235446 W001235446 W001235446 W001235446	Text Text Text A02 A02 A02 A03 A04 A05 A06 A07 A08 A09 A09 A10	Fixed Location Bar code ACD 124557875 ACD 124557877 ACD 124557877 ACD 124557877 ACD 124557879 ACD 124557879 ACD 124557880 ACD 124557881 ACD 124557881 ACD 124557882 ACD 124557881	T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244	A01 A02 A03 A04 A05 A06 A07 A08 A09 A10	Accesy	Volume 50 50 50 50 50 50 50 50 50 50 50	A22 A23 A24 A25 A26 A27 A29 A29 A30 A31	
9 1 2 3 4 5 6 7 8 9 10 11	TEXTS freezer Src IID	Text Text Text Text Text Src Poe A01 A02 A03 A04 A05 A06 A07 A06 A09 A10 A11	Fixed Location Bercode ACD 12495.7879 ACD 12455.7876 ACD 12455.7877 ACD 12455.7879 ACD 12455.7881	T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244	A01 A02 A03 A04 A05 A05 A07 A08 A09 A10 A11	Assey DMA120 DMA121 DMA121 DMA122 DMA122 DMA123 DMA125 DMA125 DMA125 DMA126 DMA127 DMA127 DMA129 DMA129 DMA129 DMA129 DMA129	Volume 30 50 50 50 50 50 50 50 50 50 5	A22 A23 A24 A25 A26 A27 A28 A29 A30 A31 A32	
9 1 2 3 4 5 6 7 8 9 10 11 12	TEXTS freezer Src IID W001235446	Text Text Text Text Text A01 A02 A03 A04 A05 A06 A07 A08 A09 A10 A11 A12	Foeed Location Bercode ACD 124557875 ACD 124557876 ACD 124557877 ACD 124557877 ACD 124557878 ACD 124557878 ACD 124557878 ACD 124557881 ACD 124557881 ACD 124557888 ACD 124557888 ACD 124557888 ACD 124557888	T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244 T003342244	A01 A02 A03 A04 A05 A06 A07 A08 A09 A10 A11 A12	ARRHY DNA120 DNA121 DNA122 DNA122 DNA123 DNA124 DNA125 DNA125 DNA125 DNA127 DNA125 DNA130 DNA131	50 50 50 50 50 50 50 50 50 50 50 50 50 5	AZ2 A23 A24 A25 A26 A27 A28 A29 A30 A31 A32 A33	

BioMicroLab XL20 with LabelPro automated label printing instrument





XL9 and XL20

Volume Detection Options

Weigh 300 vials per hour, LevelCheck 375 vials per hour

Two modules for volume detection are available with the XL20 or XL9 robotic platform: an automated balance and the LevelCheck. Both systems determine volume of tube samples stored in water, alcohol, DMSO, etc. The XL20 and XL9 can be configured with one or both technologies. The XL Series Work List Manager software includes easy-to-use data collection processing modes. For quick job set up the Tare/Gross Weighing Mode provides pre-formatted output files.

Automated Balance

Detect tube weight with a 4-place balance

- Measures 300 tubes (~3 racks) per hour
- 4-Place Analytical Balance (0.0001g readability)
- XL Ionizer is available to dissipate electrostatic charge
- · Weight data is output in .csv file
- Utilizes two rack positions on the sorting platform



Throughput Speeds For Both XL9 and XL20				
Operational Mode	Tubes / hour			
Re-array only	700			
Re-array and 2D Decode	500			
Re-array, 2D Decode, Weigh	~250			
Re-array and Weigh	~300			
Re-array, 2D Decode, LevelCheck	350-400			
Re-array, 2D Decode, DMSO Check	350-400			
Re-array and Label	180			



XL9 and XL20

Cold Room Compatibility

The XL Series accommodates lab requirements for maintaining sample temperature in cold storage, refrigeration systems and cold rooms. Cold room compatible XL Series units used for re-arraying tubes are capable of operation to -10°C. Units equipped with an automated balance are capable of operation to -4°C. Given the compact size, the XL9 and XL20 fit nicely in freezers and cold rooms where space is often limited. Cold room compatible XL Series units operate in the exact same way as they would in an ambient environment.

XL Series Work List Manager Software

User-friendly Windows based software is included with all XL Series instruments to manage sample processing modes and robotic operations. The XL Work List Manager Software provides multiple tube processing modes to meet your varied workflow needs and allows you the flexible control you need to increase your throughput. Easily managed by multiple users. Advanced system error handling options provide unattended operation. Output files can be configured to meet your LIMS system requirements.


- The user-friendly software controls sorting and re-array operations using a user generated work list (imported .csv file)
- Software generates output files for LIMS sample management database
- · Works dynamically with extended work lists
- · User is guided through the tube rack loading process
- · with a color coded interface
- Integrates with hand-held barcode scanners for rack
- identification
- · Optional: ActiveX control software

Instrument and Workflow Integration

BioMicroLab offers software developer toolkits (DLL Libraries: SDK and ActiveX) for laboratories developing in-house custom programs to control instrument operation, integrate other robotics, or streamline the processing of input and output files.

LabelPro Software

The LabelPro module includes additional capabilities built into the XL Work List Manager Software to enable you to (1) design your label with both static and merge data fields, (2) manage data to be printed, (3) preview the labels before starting a tube labeling operation, and (4) save label projects for repeated use.



Model	Capacity	Dimensions (depends on model)	Weight (depends on model) 18kg (~40 lbs) 28.3kg (~62.5 lbs) 18kg (~60 lbs)				
XL9	9 ANSI/SLAS Racks	45 x 43 x 40 cm (18" W x 17" D x 16" H)					
XL9 with LabelPro	7 ANSI/SLAS Racks	45 x 69 x 40 cm* (18" W x 27" D x 16" H*)					
XL20	20 ANSI/SLAS Racks	72 x 53 x 40 cm (28.5" W x 21" D x 16" H)					
XL20 with LabelPro	18 ANSI/SLAS Racks	72 x 79 x 40 cm* (28.5" W x 31" D x 16" H*)	37.4kg (~82.5 lbs)				
Electrical: 110-220 VAC 50/60Hz Operating Environment: 10°C to 40°C, 10-90% RH [Standard]; -4°C to 40°C, 10-90% RH [Cold Room Model] System Requirements: Windows 10, 8, 7 - 2 GB RAM Recommended - USB 2.0 ports							

